


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MM-1-C-N*
S.N.: *10749*
Date: *AUG 12 2003*



- a) Output signal amplitude:
0 to 5V (to 50Ω)
- b) Pulse width:
0.2 to 6.0 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 to 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

40dB ATTN

5 VOLT/DIV

TOP 1 NS/DIV

BOI 100 ps/DIV
(RISE TIME)

PRF = 12 MHz.