


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-P*

S.N.: *10788*

Date: *OCT 10 2003*



- a) Output signal amplitude:
0 to +5V (7050mV)
- b) Pulse width:
0.2 to 6.0ns
- c) Rise time:
≤ 100ps
- d) Fall time:
≤ 135ps
- e) PRF: *0 to 25 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

40 dB ATTEN: .5 V/DIV
770 1 ns/DIV
305 100ps/DIV (RISE TIME)
PRF = 10 MHz

+24V,
500mA (MAX)
[Signature]