


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-P*

S.N.: *10818*

Date: *DEC 08 2003*



- a) Output signal amplitude:
0 TO +5V (TO 50Ω)
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120 - 240V
50 - 60 Hz

40 dB ATN - 5V/DIV
TOP 1.0 NS/DIV
BOI 100 ps/DIV (RISE TIME)
PRF = 12 MHz