


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANL-3AH-C*

S.N.: *10849*

Date: *FEB 18 2004*



- a) Output signal amplitude:  
*0 TO +450 V (70.5uV)*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 1 NS*
- d) Fall time:  
*≤ 4 NS*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*100 - 240V*  
*50 - 60 Hz*

*60 LB ATTN*  
*- 100 V/DIV*  
*TOP 10 NS/DIV*  
*BOT 2 NS/DIV (RISE TIME)*  
*PRF = 2 KHz*