


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVIV-11A-P*

S.N.: *10882*

Date: *APRIL 23 2004*



- a) Output signal amplitude:
0 TO +50 V (TO 50Ω)
- b) Pulse width:
2 TO 100ns
- c) Rise time:
≤ 250ps
- d) Fall time:
≤ 2ns
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

50dB ATTN. ∴ 32V/DIV

TR: 5 ns/DIV

POT: 500ps/DIV (RISE TIME)

P.R.F = 10 KHz

+15 VDC

300 mA MAX