


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-P-EW-D-M*

S.N.: *10895*

Date: *APRIL 27 2004*



- a) Output signal amplitude:
0 to +5V (to 50m)
- b) Pulse width:
0.2 to 6.0 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 to 25 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

40 dB ATTN.: 5 V/DIV

TOP 1 ns/DIV

Dot 100 ps/DIV (RISE TIME)

PRF ≈ 12 MHz.

*100 → 240 V
50 - 60 Hz.*