


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-SA-C-P*

S.N.: *10899*

Date: *MAY 8 2004*



- a) Output signal amplitude:  
*0 to +20 V (70.50u)*
- b) Pulse width:  
*≤ 200ps*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 to 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*100 → 200 V*  
*50 - 60 MHz*

*40 dB ATTN. ∴ 10 V/DIV*  
*TOP: 200ps/DIV*  
*BOT: 1 μs/DIV*  
*PRF: 1 MHz*

*[Signature]*