


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVIG-1-C-N*

S.N.: *10905*

Date: *JUNE 2 2004*



- a) Output signal amplitude:  
*0 TO -5V (TO 50 mV)*
- b) FALL TIME CONSTANT:  
*5 TO 50 ns (??)*
- c) Rise time:  
*≤ 1 ns*
- d) Fall time:  
*1.47*
- e) PRF: *0 TO 20 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*30 dB ATTN.: 3.2V/DIV*  
*TIME CONSTANT FINE*  
*IN MAX CW FOR ALL*  
*THREE RANGES.*  
  
*PRF = 20 kHz.*

*100 → 240V*  
*50 \* 60 Hz*  
  
*[Signature]*