



P.O. BOX 265
OGDENSBURG, NY
U.S.A. 13669-0265
TEL: (315) 472-5270
FAX: (613) 226-2802

TEL: 1-800-265-6681
FAX: 1-800-561-1970


e-mail: info@avtechpulse.com
<http://www.avtechpulse.com/>

BOX 5120, LCD MERIVALE
OTTAWA, ONTARIO
CANADA K2C 3H4
TEL: (613) 226-5772
FAX: (613) 226-2802

PERFORMANCE CHECKSHEET

Model: AVR-3-PW-C-P-TEK2
S.N.: 10926
Date: June 9, 2004

Narrow Pulse to External 50 Ohm Load Test:




- a) Output Signal Amplitude: 0 – 250 Volts to 50 Ohms (i.e., 0 to +5 Amps)
- b) Pulse Width: 0.25 - 250 us (maximum duty cycle 0.25%)
- c) Rise Time: < 10 ns (20-80%)
- d) Fall Time: < 10 ns (80-20%)
- e) PRF: 0 - 10 kHz (maximum duty cycle 0.25%)

Output of "OUT" connector, terminated into an external 50 Ohm load. Viewed with TDS3052 scope. 50V/div, 400 ns/div.

f) Jitter, Stability: OK


Narrow Pulse to A6303 Probe and Internal Load Test:



g) Prime Power: 100-240V AC, 50-60 Hz.


Output of A6303 probe, viewed with TDS3052 scope. The A6303 probe is clamped to the shorting bar.

Wide Pulse to External 50 Ohm Load Test:



Output of "OUT" connector, terminated into an external 50 Ohm load.
Viewed with TDS3052 scope. 50V/div, 40 us/div.

Wide Pulse to A6303 Probe and Internal Load Test:



Output of A6303 probe, viewed with TDS3052 scope.
The A6303 probe is clamped to the shorting bar.

Rise, fall time reference levels: 20%, 80%.