


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-SB-T
S.N.: 10927
Date: JUNE 14 2004



- a) Output signal amplitude:
0 to +30V (to 50Ω)
- b) Pulse width:
≤ 350 ps.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 1 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

40dB ATTN.: 10V/DIV
TOP 100 ps/DIV
BOT 500 ps/DIV
PRTF = 100 KHz

+15VDC
200 mA max