


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMMH-2-C-P-ECL*

S.N.: *10944*

Date: *JULY 6 2004*



- a) Output signal amplitude:
0 TO +30V (TO 50m)
- b) Pulse width:
≤ 400ps
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 dB ATTEN
10 V/DIV

PRF = 12 MHz.

100 → 240 V
50-60 Hz.