


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AUP-AU-AU2-C-P*

S.N.: *10966*

Date: *Aug 30 2004*



- a) Output signal amplitude:
0 to +20V (to 50Ω)
- b) Pulse width:
0.3 to 2 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 200 ps
- e) PRF:
0 to 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

100 → 240V
50 - 60 Hz

[Signature]

40dB ATTEN
-10 V/DIV

TOP 1 ns/DIV

BOI 200 ps/DIV (RISE & FALL)

PRF ≈ 100 KHz.