


PULSE GENERATOR

PERFORMANCE CHECK

Model: ANM14-5-C-F

S.N.: 10711

Date: OCT 7 2003



- a) Output signal amplitude:  
0 TO +100V (TO 50V)
- b) Pulse width:  
2 TO 4NS
- c) Rise time:  
≈ 1 NS
- d) Fall time:  
≈ 2 NS
- e) PRF:  
0 TO 1 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240V  
50-60 Hz

60 LB ATTEN

100 V/DIV

PRF = 100 KHz

*[Handwritten signature]*