


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVD*

S.N.: *856*

Date: *APRIL 30 82*



- a) Output signal amplitude:
+10 TO +50 AMPS
- b) Pulse width:
20 ± 5 nSEC.
- c) Rise time: (HALF POWER)
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

*25 AMP/DIV
10 nSEC/DIV
1 KHZ.*

*+15 ± 1.0 VOLT.
150 mA (MAX)*