


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-1-P.A-PS

S.N.: 1001

Date: AUG 6 82



a) Output signal amplitude:
0 to +250 V

b) Pulse width:
≤ 4 nsec

c) Jitter, stability:
OK

d) Prime power:
120/240 V/
50-60 Hz

100 V/DIV.
2 nsec/DIV.
40 kHz

NOTE: UNIT
SHIPPED
WITH VOLTAGE
SELECTOR
CIRCUIT IN
120 V
POSITION.