


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2A-W-P-PS*

S.N.: *1018*

Date: *JAN 10 83*



*≈ 100 VOLTS/DIV
10 NSEC/DIV
4 KHz*

a) Output signal amplitude:

0 TO 7150V

b) Pulse width:

*0 TO 400 NSEC**

c) Rise time:

≈ 1 NSEC

d) Fall time:

≈ 2 NSEC

e) PRF:

0 TO 5 KHz

f) Jitter, stability:

OK

g) Prime power:

110/220V

50-60 Hz

[Signature]

** REQUIRES ADDITION
OF EXTERNAL DELAY
LINE FROM 100 TO
400 NSEC RANGE.*