


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANP-AU-1-P-C*

S.N.: *1020*

Date: *SEPT 27 82*



10 VOLTS/DIV.
1.0 NSEC/DIV.
400 KHz.

- a) Output signal amplitude:
0 TO 410 V
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:†
0 TO 1 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

110/220 V
50-60 Hz

** 1 KHz TO*
1 MHz ON
INT CLOCK.