


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-S-1-P*

S.N.: *1039*

Date: *AUG 22 82*



) Output signal amplitude:

*> +10V*

) Pulse width:

*≈ 130 psec*

) Jitter, stability:

*OK*

) Prime power:

*+15 ± 1.0V*

*170 mA (max)*

*10 VOLTS/DIV*

*1.0 NSEC/DIV*

*400 KHz*