

IMPULSE GENERATOR

PERFORMANCE CHECK

Model:


AVH-HV1-C

S.N.:

1052

Date:

DEC 20 82



a) Output signal amplitude:

$\geq +100V$ TO 50Ω

b) Pulse width:

≤ 1 NSEC

c) Jitter, stability:

OK

d) Prime power:

110/220 V 50-60 Hz.

e) DELAY:

- 125 NSEC TO

+ 8.5 msec.

≈ 32 VOLTS/DIV

1.0 NSEC/DIV

100 KHz.