


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AP7-AU-1411-C-M*

S.N.: *1063*

Date: *NOV 30 82*



(A) *10 VOLTS / DIV*
1.0 NS / DIV
400 KHz



(B) *AS (A) BUT*
MONITOR
OUTPUT

- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
0 TO 2 NS
- c) Rise time:
≤ 100 PSEC.
- d) Fall time:
≤ 150 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK.
- g) Prime power:
110/220V
50 - 60 Hz