


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-IMPAT-N*

S.N.: *1089*

Date: *DEC 17 82*



a) Output signal amplitude:

*7-32 VOLTS*

b) Pulse width:

*≤ 1 nsec*

c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 1.0V*

*110 mA (max)*

*≈ 32 VOLTS/DIV.*

*1.0 nsec/DIV.*

*400 kHz*