


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-5-1-W-A

S.N.: 1130

Date: NOV 30 82



a) Output signal amplitude:

0 to 10V

b) Pulse width:

≤ 130 PSEC

c) Jitter, stability:

OK.

d) Prime power:

+15 ± 1.0 V

175 mA ~~mA~~

10 V / DIV  
1.0 NSEC / DIV  
400 KHZ.

Handwritten signature or initials, possibly 'AJ'.