


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-P-P5*
S.N.: *1157*
Date: *MAR 7 83*



- a) Output signal amplitude:
0 TO +35.0 V
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≤ 1.5 NSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*≈ 320 V / DIV
10 NSEC / DIV
2 KHz*

*120 / 240 V
50 - 60 Hz*