


PULSE GENERATOR

PERFORMANCE CHECK


Model: *APV-AU-2-N-05-C*

S.N.: *1159*

Date: *DEC 30 82*



*2 nSEC/DIV  
10 VOLTS/DIV  
40 KHz.*



*10 nSEC/DIV  
10 VOLTS/DIV  
40 KHz.*

- a) Output signal amplitude:  
*0 TO -10V*
- b) Pulse width:  
*5 TO 50 nSEC*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 135 pSEC*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*110/220V  
50-60 Hz*