


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVI 41*
S.N.: *1167*
Date: *FEB 9 83*



- a) Output signal amplitude:
71 +40V
- b) Pulse width:
1 TO 100 NSER
- c) Rise time:
≤ 300 PSEC
- d) Fall time:
≤ 2 NSER
- e) PRF:
SEE FIG 1 & 2
- f) Jitter, stability:
OK
- g) Prime power:

+15V ± 1.0V
70 mA (MAX) *✓*

1 NSER/DIV
32 VOLTS/DIV
40 KHZ