


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PAVE-41*
S.N.: *1170*
Date: *FEB 9 83*



- a) Output signal amplitude:
> 40 V
- b) Pulse width:
1 TO 100 nSEC
- c) Rise time:
≤ 300 PSEC
- d) Fall time:
≤ 2 nSEC
- e) PRF:
SEE T-16 1-12
- f) Jitter, stability:
OK
- g) Prime power:

*32 VOLTS/DIV
1 NSEC/DIV
40 KHz.*

*+15 ± 1.0 V
70 mA (MAX)
*[Signature]**