


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-P*

S.N.: *1204*

Date: *MAR 17 83*



- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
0 TO 2 μ SEC
- c) Rise time:
 \leq 100 PSEC
- d) Fall time:
 \leq 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV.
1.0 μ SEC/DIV.
4.0 MHz:*

*+24 \pm 1.0V
500 mA (MAX)*