


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-3-P*  
S.N.: *1205*  
Date: *APRIL 18 83*



*10 VOLTS/DIV.*  
*2.0 NSEC/DIV.*  
*4.0 MHz.*

- a) Output signal amplitude:  
*0 TO +15V*
- b) Pulse width:  
*2.0 TO 15 NSEC*
- c) Rise time:  
*< 150 PSEC*
- d) Fall time:  
*≤ 600 PSEC.*
- e) PRF:  
*0 TO 25 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+24 TO +30 VOLTS,*  
*450 mA (MAX)*