

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVL-2-P

S.N.:

1230

Date:

FEB 15 83



- a) Output signal amplitude:
0 to +350V
- b) Pulse width:
0 to 100 nsec
- c) Rise time:
1.5 nsec.
- d) Fall time:
2 nsec
- e) PRF:
0 to 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:

320 VOLTS/DIV
20 nsec/DIV
2 kHz.

+15 ± 1.0 V
500 mA (MAX)