


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-V-A-P-PS*

S.N.: *1254*

Date: *FEB 28 83*



- a) Output signal amplitude:
0 TO +50 V
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≤ 500 PSEC
- d) Fall time:
≤ 2.0 NSEC
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*120/240 V
50 - 60 kHz*

*32 VOLTS / DIV
5 NSEC / DIV
10 KHz*

[Signature]