


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-0-2-P-TTL*

S.N.: *1260*

Date: *MAY 10 83*



a) Output signal amplitude:

*+10V*

b) Pulse width:

*2 TO 50 nSEC*

c) Rise time:

*≤ 100 pSEC*

d) Fall time:

*≤ 135 pSEC*

e) PRF:

*0 TO 50 KHz.*

f) Jitter, stability:

*OK*

g) Prime power:

*+15 ± 1.0 V*

*250 mA (max)*

*10 VOLTS/DIV.*

*2.0 nSEC/DIV*

*40 KHz.*