


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVG-1-P*

S.N.: *1265*

Date: *APRIL 11 83*



a) Output signal amplitude:

0 TO + 250 V

b) Pulse width:

≤ 4 nSEC

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 V
400 mA (MAX)

100 VOLTS/DIV
5 nSEC/DIV
40 kHz