


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-2-N-D*

S.N.: *1292*

Date: *MAR 17 83*



a) Output signal amplitude:

0 TO -15 V

b) Pulse width:

0 TO 2 NSEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 ± 1.0 V

500 mA (MAX)

10 VOLTS/DIV
1.00 NSEC/DIV
7.0 MHz.