


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-2-PS-P*

S.N.: *1299*

Date: *APRIL 29 83*



- a) Output signal amplitude:  
*0 TO +350 V*
- b) Pulse width:  
*0 TO 100 NSEC*
- c) Rise time:  
*≤ 1.5 NSEC*
- d) Fall time:  
*≤ 3 NSEC*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*110/220 V*  
*50 - 60 HZ*

*± 320 VOLTS/DIV.*

*10 NSEC/DIV.*

*2 KHZ*

*[Signature]*