


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVP-AV-2-N-05-PS

S.N.: 1308

Date: MAY 31 1963



- a) Output signal amplitude:  
0 to 10V
- b) Pulse width:  
5 to 50 nsec
- c) Rise time:  
≤ 100 psec
- d) Fall time:  
≤ 135 psec
- e) PRF:  
0 to 50 kHz
- f) Jitter, stability:  
OK
- g) Prime power:

10 VOLTS/DIV  
2 nsec/DIV  
20 kHz

110/220V  
50-60 Hz

*[Signature]*