

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AV~~A~~-AV-1-N-TTL-05-PS

S.N.:

1315

Date:

APRIL 25 1983



- a) Output signal amplitude:
0 TO 10 V
- b) Pulse width:
0 TO 4 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
1.0 NS/DIV
400 KHz

120/240 V
50-60 Hz
UNIT SHIPPED
WITH VOLTAGE
SELECTOR CARD
IN THE 240 V
POSITION.