


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-5-1-C-N

S.N.: 1374

Date: MAY 24 83



a) Output signal amplitude:

0.75 - 1.0 V

b) Pulse width:

≤ 135 psec

c) Jitter, stability:

OK

d) Prime power:

110/220 V

50-60 Hz.

[Handwritten signature]

1.0 nsec/div
10 volts/div
400 kHz.