


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVD-M-1-D*

S.N.: *1387*

Date: *AM 24 83*



- a) Output signal amplitude:
0 TO + 10 V
- b) Pulse width:
0 TO 4 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
1.0 NS/DIV
400 RATE

+15 ± 1.0V
275 mA (MAX)