


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PVL-2A-W-P-PS*

S.N.: *1391*

Date: *JULY 21 83*



*100 V/DIV
10 nsec/DIV
2 kHz*

- a) Output signal amplitude:
0 TO +170 V
- b) Pulse width: ***
0 TO 400 nsec
- c) Rise time:
≤ 1 nsec
- d) Fall time:
≤ 2 nsec
- e) PRF:
0 TO 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:

*110/220 V
50-60 Hz.*

** REQUIRES EXTERNAL
DELAY LINE FOR
PW 100 TO 400 nsec.*