


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ALM-2-N-M*

S.N.: *1393*

Date: *JUNE 27 83*



*10 VOLTS/DIV  
100 NSEC/DIV  
4.0 MHz.*

- a) Output signal amplitude:  
*0 TO -15 V*
- b) Pulse width:  
*0 TO 2 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+24 ± 1.0 VOLT;  
500 mA.*
- h) MONITOR OUTPUT:  
*0 TO 1.5 V  
(20 db DOWN WRT  
MAIN OUTPUT).*

*[Signature]*