


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-P*

S.N.: *1394*

Date: *Aug 30 83*



- a) Output signal amplitude:
10 TO 50 AMP)
- b) Pulse width:
23 nsec
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*20 Amps / DIV.
10 nsec / DIV
1 KHz.*

*+15 ± 1.0 V
175 mA (max)*