


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-5A-C-D*

S.N.: *1399*

Date: ~~SEP~~ *SEPT 7 83*



- a) Output signal amplitude:  
*0 TO +40 AMP*
- b) Pulse width:  
*20 TO 150 nSEC*
- c) Rise time:  
*≤ 5 nSEC*
- d) Fall time:  
*≤ 5 nSEC*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*≈ 40 AMPS / DIV*  
*10 nSEC / DIV*  
*5 KHz*

*(OSC DUE TO PROBE)*

*120 / 240 V*  
*50 - 60 Hz*

*UNIT SHAPERS WITH VARIABLE SELECTOR KNOB IN THE 240V POSITION*

*M*