


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVL-TN-1*  
S.N.: *1416*  
Date: *JUNE 23 83*



*100 V/DIV*  
*10 NS/DIV*  
*2 kHz*

- a) Output signal amplitude:  
*-230V (+30V OFFSET)*
- b) Pulse width:  
*0 TO 100 NS*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 3 NS*
- e) PRF:  
*0 TO 5 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*110/220V*  
*50-60 Hz*

*t<sub>p</sub> : ≤ 100 NS*