


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-1-N*

S.N.: *1420*

Date: *JUNE 17 83*



- a) Output signal amplitude:
0 TO 5V
- b) Pulse width:
0 TO 6 NSCZ
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MITE
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV
1.0 NSCZ/DIV
4.0 MITE.*

~~*24*~~
*+24 ± 1.0 VOLT
400 mA (MAX)*