


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PTM-2-N*
S.N.: *1425*
Date: *June 13 83*



- a) Output signal amplitude:
0 TO -15V
- b) Pulse width:
0 TO 2 μ SEC
- c) Rise time:
 ≤ 100 μ SEC
- d) Fall time:
 ≤ 135 μ SEC
- e) PRF:
0 TO 25 μ Hz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
1.0 μ SEC/DIV
4.0 μ Hz

+24 \pm 1.0 V
500 mA (MAX)
[Signature]