


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-TN-1
S.N.: 1450
Date: JUNE 23 83



- a) Output signal amplitude:
-230V (+30V OFFSET)
- b) Pulse width:
0 TO 100 nsec
- c) Rise time:
< 2 nsec
- d) Fall time:
< 3 nsec
- e) PRF:
0 TO 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:
110/220V
50-60 Hz

100V/DIV.
10 nsec/DIV
2 kHz

BP: < 100 nsec