


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-2-C-D-N*

S.N.: *1452*

Date: *Nov 29 83*



a) Output signal amplitude:

*0 TO -15 V*

b) Pulse width:

*0 TO 2 NS*

c) Rise time:

*≤ 100 PSEC*

d) Fall time:

*≤ 135 PSEC*

e) PRF:

*0 TO 25 MHz*

f) Jitter, stability:

*OK.*

g) Prime power:

*110/220 V*

*50-60 MHz*

*10 V/DIV*

*1.0 NS/DIV*

*4.0 MHz*