


PULSE GENERATOR

PERFORMANCE CHECK

Model: A20-1  
S.N.: 1477  
Date: Jul 30 83



- a) Output signal amplitude: 25 TO 100 AMP
- b) Pulse width: 25 NSEC
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 1 KHZ
- f) Jitter, stability: OK
- g) Prime power:

60 AMPS/DIV  
10 NSEC/DIV  
1 KHZ

+15 ± 1.0 V  
190 mA (MAX)