



PULSE GENERATOR
PERFORMANCE CHECK

Model: AUP-AV-2-F-C-05-M-W
(RECALIBRATED, JAN 25)
S.N.: 1479
Date: JAN 25 85



MAIN OUTPUT
10 VOLTS/DIV
1.0 NSEC/DIV



MONITOR OUTPUT
1 VOLT/DIV
1.0 NSEC/DIV

- a) Output signal amplitude:
- b) Pulse width:
- c) Rise time:
- d) Fall time:
- e) PRF:
- f) Jitter, stability:
- g) Prime power:

} THIS DEGREE OF
OVERSHOOT IS TYPICAL
FOR OUR MODELS.




PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVP-AV-2-P-C-05-M-W*

S.N.: *1479*

Date: *OCT 3 83.*



① RISE TIME
1 NSER/DIV
10 VOLTS/DIV
25 KHZ



② FALL TIME
1 NSER/DIV
10 VOLTS/DIV
25 KHZ

a) Output signal amplitude:

0 TO +15V

b) Pulse width:

20 TO 500 NSER

c) Rise time:

≤ 150 PSEZ

d) Fall time:

≤ 200 PSEZ

e) PRF:

0 TO 30 KHZ

f) Jitter, stability:

OK

g) Prime power:

110/220V

50-60 Hz.