


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-MP-1*

S.N.: *1486*

Date: *SEPT 26 83*



- a) Output signal amplitude:  
*7, ± 40V*
- b) Pulse width:  
*1 TO 100 NSEC*
- c) Rise time:  
*< 500 PSEC*
- d) Fall time:  
*< 2 NSEC*
- e) PRF:  
*SEE FILE 1+2*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*3.2 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*40 KHZ*

*± 15% 1.0*  
*150 mA (MAX)*  
*dy*